



California Medical Device Recall Information



Recall Name

**Philips Respironics Recalls Trilogy Ventilator Models 100, 200 and 202
Due to a Potential Defect**

Recall Date	Product Description	Recalling Firm	Recall Reason
2/11/14	Philips Respironics Trilogy Ventilators Trilogy Ventilator Models 100, 200 and 202	Respironics, Inc. Murrysville, PA	<i>A potentially defective component on the power management board could cause the ventilator to fail to deliver mechanical breaths.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Models 100, 200 and 202 Product Labels	CA , globally	Distributed between: December 31, 2013 and January 30, 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm386490.htm>