



California Medical Device Recall Information



Recall Name

Quest Medical Recalls the Myocardial Protection System (MPS) Delivery Sets Due to a Potential for Seal Failure

Recall Date	Product Description	Recalling Firm	Recall Reason
10/28/15	Myocardial Protection System (MPS) Delivery Sets	Quest Medical, Inc. Allen, TX	<i>Potential for seal failure during use, which could result in patient blood loss.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Models: <ul style="list-style-type: none"> • 5001102 • 5001102-AS • 7001102 List of Suspect S/Ns	CA, nationwide	Manufactured from: May 2015 to September 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm470146.htm>