



California Medical Device Recall Information



Recall Name

**Richard-Allan Scientific Recalls 10% Neutral Buffered Formalin
Due to Potential for Incorrect Concentration of Formalin**

Recall Date	Product Description				Recalling Firm	Recall Reason
09/17/14	10% Neutral Buffered Formalin <i>Affected products are in table below</i>				Richard-Allan Scientific Kalamazoo, MI	<i>Potentially incorrect concentration of Formalin (0% - 3%).</i>
Recall Class	Product Identification				Distribution	Affected Dates
I	Product Code	Product Name	Lot Number	Exp. Date	CA, nationwide	Distributed: July 18, 2014 to Sept. 17, 2014
	534801	Richard-Allan Scientific Specimen Transportation System	296986	July 2018		
	59601	Richard-Allan Scientific Bio-Tite Specimen Containers	300233	July 2018		
	51401	Richard-Allan Scientific 10% Neutral Buffered Formalin	300649	June 2018		
	53151	Richard-Allan Scientific Specimen Transportation System	300143	July 2018		
	426-796	Protocol 10% NBF	298987	July 2018		
	C4320-450B	Cardinal Health Prefilled Buffered 10% Formalin	300403	July 2018		
	LC-0060	Richard-Allan Scientific 10% Neutral Buffered Formalin	300445	July 2018		

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm421411.htm>