



California Medical Device Recall Information



Recall Name

**Respironics California Recalls Esprit V1000 and V200 Ventilators, and
3rd Generation Power Supply Repair Part (RP) Kits
Due to Potential for Power Failure**

Recall Date	Product Description	Recalling Firm	Recall Reason
09/17/14	Esprit Ventilators <ul style="list-style-type: none"> Model V1000 Model V200 3rd Generation Power Supply Repair Part (RP) Kits	Respironics California, Inc. Carlsbad, CA	<i>Faulty 3rd Generation Power Supply may prevent the ventilator from using AC power (electricity from a wall socket), or may fail and prevent the ventilator from switching back to AC power after using battery power.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Affected ventilator models were installed with 3rd Generation Power Supplies; 3rd Generation Power Supply Repair Part (RP) Kits List of Affected Serial Numbers	CA, nationwide	Manufactured and distributed from: December 21, 2012 to July 9, 2014.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm423848.htm>