



California Medical Device Recall Information



Recall Name

Symbios Medical Products Recalls GOPump Elastomeric Infusion Pump Kit Due to a Possible Malfunction

Recall Date	Product Description	Recalling Firm	Recall Reason
2/25/13	GOPump Elastomeric Infusion Pump Kit	Symbios Medical Products, LLC. Indianapolis, IN	<i>Potential for solutions to flow at a higher rate than intended.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	GOPump Elastomeric Infusion Pump Kit (with and without accessories) containing: <ul style="list-style-type: none">Dual GOPump [Part Number: 510076] List of Suspected Lots Recalled	CA, nationwide	Distributed between September 10, 2012 and February 11, 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm344311.htm>