



# California Medical Device Recall Information



## Recall Name

### Teleflex Medical Recalls ISIS HVT Tracheal Tube Due to Potential Kinking During Patient Use

Recall Date	Product Description	Recalling Firm	Recall Reason
01/06/14	ISIS HVT Tracheal Tube Cuffed with Subglottic Secretion Port (with and without Preloaded Stylet)	<b>Teleflex Medical</b> Research Triangle Park, NC	<i>The affected tracheal tube may kink during patient use.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Sizes: 6.0 – 9.0 <a href="#">Suspected List of Affected Product Codes and Lot Numbers</a>	CA, nationwide	Distributed from: March 2010 through December 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm386984.htm>