



California Drug Recall Information



Recall Name

**Wallcur, Inc. Recalls Simulated Intravenous (IV) Saline Solution
Due to Non-Sterile Products Being Administered to Patients**

Recall Date	Product Description	Recalling Firm	Recall Reason
01/07/15	<ul style="list-style-type: none"> Practi-0.9% sodium chloride IV bags; 50 mL, 250 mL, 500, mL, and 1000 mL sizes Practi-0.9% sodium chloride IV bag with sterile distilled water; 100 mL size 	Wallcur, LLC San Diego, CA	<p><i>Use of these products on humans or animals could result in serious adverse events.</i></p> <p><i>Wallcur products are intended for training, simulation, and educational purposes only and are not sterile.</i></p> <p><i>Adverse events have been reported after accidental administration to patients.</i></p>
Recall Class	Product Identification	Distribution	Affected Dates
N/A	Product Photos	CA, nationwide	Products shipped on and after: May 22, 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm429724.htm>