



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Phillips Respironics CPAP, BiPAP, and Ventilator Recall

Recall Date	Product Description	Recalling Firm	Recall Reason
06/14/2021	<p>Continuous Ventilator, Minimum Ventilatory Support, Facility Use E30 Emergency Use Authorization</p> <p>A-Series BiPAP V30 Auto Ventilator Trilogy 100, 200</p> <p>All serial numbers Manufactured before 4/26/21</p>	<p>Phillips Respironics, Murrysville, PA</p>	<p>Sound Abatement Foam (PE-PUR) susceptible to degradation and particle release, exacerbated by ozone cleaning.</p>
06/14/2021	<p>Continuous Ventilator, Non-life Supporting DreamStation ASV BiPAP autoSV DreamStation ST, AVAPS SystemOne ASV4 BiPAP autoSV, BiPAP autoSV Advanced C-Series ASV C-Series S/T and AVAPS OmniLab</p> <p>All serial numbers Manufactured before 4/26/21</p>	<p>Phillips Respironics, Murrysville, PA</p>	<p>Sound Abatement Foam (PE-PUR) susceptible to degradation and particle release, exacerbated by ozone cleaning.</p>
06/14/2021	<p>Noncontinuous Ventilator SystemOne (50,60-Series)CPAP, Auto CPAP, BiPAP DreamStation CPAP, Auto CPAP, BiPAP DreamStation Go CPAP, APAP, Auto CPAP All serial numbers Manufactured before 4/26/21</p>	<p>Phillips Respironics, Murrysville, PA</p>	<p>Sound Abatement Foam (PE-PUR) susceptible to degradation and particle release, exacerbated by ozone cleaning.</p>

Recall Class	Product Identification	Distribution	Affected Dates
I	Dreamstation series All serial numbers	Nationwide, including California	April 26, 2021 and earlier
Voluntary	E30, A-Series, Trilogy, OmniLab, C-Series, SystemOne All Serial Numbers	Nationwide, including California	April 26, 2021 and earlier

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [PHILLIPS WEBSITE](#) or [FDA](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

