



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Baxter Healthcare Corporation Recalls Volara System For Risk Of Respiratory Distress In Ventilated Patients During Home Use

Recall Date	Product Description	Recalling Firm	Recall Reason
4/26/2022	In-Line ventilator adaptor M08473 OPTIMUS OLE	Baxter Healthcare Corporation Deerfield, Illinois	There is a potential of reduced oxygenation or pneumothorax/barotrauma to occur when the Volara system is used with in-line ventilator in home care environment.
4/26/2022	In-Line ventilator adaptor M07937 MODULE, OPTIMUS HANDSET 2	Baxter Healthcare Corporation Deerfield, Illinois	There is a potential of reduced oxygenation

Recall Class	Product Identification	Distribution	Affected Dates
I	In-Line ventilator adaptor M08473 OPTIMUS OLE AC PAT.CIRCUIT KIT; UDI-DI: 00887761981492.	259 Units Nationwide including California	February 2020 to April 2022
I	In-Line ventilator adaptor M07937 MODULE, OPTIMUS HANDSET 2; UDI-DI: 00887761984622.	9 Units Nationwide including California	October 2019 to April 2022

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

