

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Alcon Research LLC Monarch ζ III D Delivery Cartridge

Recall Date	Product Description	Recalling Firm	Recall Reason
01/10/2025	<p>Brand Name: Alcon</p> <p>Product Name: Monarchζ III D Delivery Cartridge</p> <p>Model/Catalog Number: 8065977763</p> <p>Product Description: Monarchζ III D delivery cartridge Lens Guide</p> <p>Intraocular Component: N/A</p>	Alcon Research LLC	Increase in complaints associated with a molding defect within the inner lumen of delivery cartridges which could result in a scratch on the intraocular lens (IOL) upon delivery.

Recall Class	Product Identification	Distribution	Affected Dates
II	<p>GTIN/UDI 10380659777636</p> <p>Lot # 15895927 - 15950096</p>	2,134 units in California	November and prior

For additional information, please visit the [FDA Website](#).