State of California—Health and Human Services Agency

California Department of Public Health

Food and Drug Branch – Device Recalls

California Device Recall Information Sheet

Ambu Recalls Spur Ii Adult Resuscitator

Recall Date	Product Description	Recalling Firm	Recall Reason
8/20/2025	**Ambu Spur Ii** Labeled as the following: 1 SPUR II Adult Resuscitator, Adult, Catalog Number: 523211000 2 SPUR II Adult Resuscitator w/Mercury Filter …	**AMBU INC.** Columbia, Maryland	Potential for the manometer port being blocked rendering the manometer non-functional.

Recall Class	Product Identification	Distribution	Affected Dates
I	**Ambu Spur II** 1. Catalog Number: 523211000; UDI Number: 05707480152193; Lot Numbers: 1001113554, 1001118764. 2. Catalog Number: 523611051; UDI Number: 05707480149513 …	87156 Units Worldwide, including US	August 2025 and prior

For additional information, please visit the FDA website

CDPH Food and Drug Branch
MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435
(916) 650-6500 ● (916) 650-6650 FAX
Internet Address: www.cdph.ca.gov