California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

CareFusion 303, Inc. – BD Pyxis MedStation ES and Pyxis Pro Automated Dispensing Cabinets Recall

Recall Date	Product Description	Recalling Firm	Recall Reason
12/19/2025	BD Pyxis MedStation ES and Pyxis Pro Automated Dispensing Cabinets are medication storage and dispensing systems used in healthcare settings to securely manage and distribute medications.	CareFusion 303, Inc. San Diego, CA, USA	A firmware update that may cause duplicate address errors in CUBIE drawers, leading to drawer failure and potential delays in accessing medications.

Recall Class	Product Identification	Distribution	Affected Dates
II	**Lot Code:** N/A **Model No.:** 323, 324, 352, 1149-00, 1152-00, 1155-00, 1116-00 **UDI-DI:** **323**: 10885403512667 **324**: 10885403512681 **352**: 10885403512674 **1149-00**: 10885403517822 **1152-00**: 10885403517839 **1155-00**: 10885403521300 **1116-00**: 10885403512605 **S/N range:** N/A **Firmware Version**: 1.48	53,669 Units; Nationwide including California	November 2025 and prior (firmware version 1.48)

For additional information, please visit the FDA Website.