California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

CareFusion 303, Inc. – BD Pyxis ES Enterprise Server Recall

Recall Date	Product Description	Recalling Firm	Recall Reason
01/08/2026	BD Pyxis ES Enterprise Server is a web-based application used to configure and manage BD Pyxis automated dispensing cabinets and associated medication management systems.	CareFusion 303, Inc. San Diego, CA, USA	Software issue may delay updates of patient, medication, or order information, potentially leading to delayed or incorrect clinical decisions.

Recall Class	Product Identification	Distribution	Affected Dates
II	**Lot Code:** N/A **Model No.:** Multiple catalog numbers including 1115-00, 137672-02/03, 137671-02/03, 137670-02/03, and more (see full list in recall data) **UDI-DI:** **Hardware**: 10885403518348 **Software**: 10885403520341 **S/N range:** N/A **Software Versions**: System Versions 1.6.1 – 1.11.0; Server Application Versions ES 4.13 – 5.8.X	2,079 Systems; Nationwide including California	December 2025 and prior.

For additional information, please visit the FDA Website.