

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

BD Switzerland Sarl BD Connecta BD Luer-Lok 360

Recall Date	Product Description	Recalling Firm	Recall Reason
04/02/2025	BD Connecta BD Luer-Lok 360 REF: 394910 UDI-DI: 00382903949106 Connecta Plus 1 and Connecta Plus 3 2-way and 3-way Stopcocks are control valves for use in IV. therapy and hemodynamic pressure monitoring.	BD Switzerland Sarl	Due to specific lots of luer-lok devices were shipped to the U.S. market with an Outside US Instruction for Use (IFU)

Recall Class	Product Identification	Distribution	Affected Dates
II	Catalog Number: 394910 UDI-DI: 00382903949106 Lot Numbers: 4277968 - 4306700	500 units nationwide	February and prior

For additional information, please visit the [FDA Website](https://www.fda.gov).