

## California Device Recall Information Sheet

### Food and Drug Branch – Device Recalls

Baxter Healthcare Corporation Novum IQ Large Value Pump Infusion System

Recall Date	Product Description	Recalling Firm	Recall Reason
05/20/2025	Baxter Novum IQ Large Value Pump Infusion System  REF: 40700BAXUS	Baxter Healthcare Corporation	Due to the potential for underinfusion following use of the "standby mode" feature or if the device is powered off with the set loaded.

Recall Class	Product Identification	Distribution	Affected Dates
I	All serial numbers	34,524 units nationwide	April and prior

For additional information, please visit the [FDA Website](#).