California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Baxter Healthcare Corporation Breathe Technologies, Inc., and Welch Allyn, Inc., Life2000 Ventilation System

Recall Date	Product Description	Recalling Firm	Recall Reason
01/16/2025	Breathe Technologies, Inc., Life2000 Ventilation System, REF BT-20-0002AP Welch Allyn, Inc., Life 2000 Ventilation System, REF BT-20-0002	Baxter Healthcare Corporation	There is an issue with the ventilator battery charger that could lead to an audible and visual alarm. If the alarm is engaged, the ventilator becomes inoperable.

Recall Class	Product Identification	Distribution	Affected Dates
I	UDI/DI: 00887761978089 or 00815410020537 Serial Numbers: 122010001081 UDI/DI 00887761978201 Serial Numbers: 124080000663 - 124280000741	1 unit nationwide 187 units nationwide	December and prior

For additional information, please visit: FDA Website (Breathe Tech.) and FDA Website (Welch Allyn).