California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Becton Dickinson & Company Recalls BDSpinal Tray

Recall Date	Product Description	Recalling Firm	Recall Reason
5/29/2026	**Product 1** BDSpinal Tray with BDWhitacre Needle 22 G x 3.5 in.; lidocaine HCL (1%), 5 mL, bupivacaine HCL (0.75%) with dextrose (8.25%), 2 mL (marcaine). Catalog Number: 405645	Becton Dickinson & Company Franklin Lakes, NJ	This is a sub-recall of a product recall from Huons Co, Ltd., the manufacturer of Bupivacaine Hydrochloride in Dextrose Injection, USP. The recall is due to quality issues identified during an FDA manufacturing facility inspection, citing specific quality deficiencies related to acceptance testing.
5/29/2026	**Product 2** BDSpinal Tray with BDQuincke Needle 26 G x 3.5 in.; lidocaine HCL (1%), 5 mL, bupivacaine HCL (0.75%) with dextrose (8.25%), 2 mL (marcaine). Catalog Number: 405632	Becton Dickinson & Company Franklin Lakes, NJ	This is a sub-recall of a product recall from Huons Co, Ltd., the manufacturer of Bupivacaine Hydrochloride in Dextrose Injection, USP. The recall is due to quality issues identified during an FDA manufacturing facility inspection, citing specific quality deficiencies related

			to acceptance testing.
5/29/2026	**Product 3** BDSpinal Tray with BDWhitacre Needle 25 G x 3.5 in.; lidocaine HCL (1%), 5 mL, bupivacaine HCL (0.75%) with dextrose (8.25%), 2 mL (marcaine). Catalog Numbers: 405637, 405707	Becton Dickinson & Company Franklin Lakes, NJ	This is a sub-recall of a product recall from Huons Co, Ltd., the manufacturer of Bupivacaine Hydrochloride in Dextrose Injection, USP. The recall is due to quality issues identified during an FDA manufacturing facility inspection, citing specific quality deficiencies related to acceptance testing.

Recall Class	Product Identification	Distribution	Affected Dates
I	**Product 1** Catalog Number: 405645. UDI Numbers: (01)00382904056452(17)261001(10)B01V223D. Lot Number (Expiration Date): B01V223D (2026-10-01)	330 Units Nationwide	April 2026 and prior
I	**Product 2** Catalog Number: 405632. UDI Number: (01)00382904056322(17)270101(10)B01V231D; Lot Number (Expiration Date): B01V231D (2027-01-01). Expansion 07/08/2026 UDI Number: (01)00382904056322(17)270101	890 Units Nationwide California	April 2026 and prior

	(10)B02V248D; Lot Number (Expiration Date): B02V248D (2027-01-01).		
I	**Product 3** Catalog Number: 405637. UDI Numbers: (1)00382904056377(17)260628(10)B01V160D, (01)00382904056377(17)270101(10)B01V241D, (01)00382904056377(17)270101(10)B01V265D, (01)00382904056377(17)261001(10)B02V183D. Lot Numbers (Expiration Date): B01V160D (2026-06-28), B01V241D (2027-01-01), B01V265D (2027-01-01), B02V183D (2026-10-01) Catalog Number: 405707. UDI Number: (1)00382904057077(17)270101(10)B02V230D. Lot Number (Expiration Date): B02V230D (2027-01-01)	5,940 Units Nationwide	April 2026 and prior

For additional information, please visit Product 1 , Product 2 , Product 3