

# California Device Recall Information Sheet

## Food and Drug Branch – Device Recalls

Becton Dickinson & Co. BD BBL TB Stain Kit K

Recall Date	Product Description	Recalling Firm	Recall Reason
03/06/2025	BD BBL TB Stain Kit K  Catalog (REF): 212522	Becton Dickinson & Co.	One component of the kit was shipped with an expiration date of 30Sep2024 while the Certificate of Analysis (COA) indicated an expiration date of 30Apr2025.

Recall Class	Product Identification	Distribution	Affected Dates
II	Catalog (REF): 212522  UDI-DI: 00382902125228  Expiration Date: 2025/04/30  Date of Manufacture: 07/10/2024  Lot Number: 4191205	247 units nationwide	February and prior

For additional information, please visit the [FDA Website](#).