



Food and Drug Branch – Drug Recalls

**CALIFORNIA DRUG RECALL INFORMATION SHEET**

**Biocheck Recalls T4 EIA Kit for Not Meeting Product Shelflife Standards**

Recall Date	Product Description	Recalling Firm	Recall Reason
11/17/2021	<b>T4 EIA Kit</b> Cat No: BC-1007, Store at 2-8C, For in vitro diagnostic use only This products intended use is For the quantitative determination of total thyroxine (T4) concentration in human serum. The test is useful in the diagnosis and treatment of thyroid disorders	<b>Biocheck Inc.</b> South San Francisco, California	Immunoassay TEST Kit do not meet product shelf-life standards resulting in failed controls.

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>T4 EIA Kit</b> Lot number: RN-60831	14 Kits in California	October 2021 and prior
II	<b>T4 EIA Kit</b> Lot numbers: RN-60923	8 Kits in California	October 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

