

California Device Recall Information Sheet

[Food and Drug Branch – Device Recalls](#)

B Braun Medical Inc Introcan Safety 3 Closed IV Catheter

Recall Date	Product Description	Recalling Firm	Recall Reason
03/19/2025	Introcan Safety 3 Closed IV Catheter 22 Ga. x 1 in., PUR, Winged Passive safety IV catheter	B Braun Medical Inc	Increased risk for valve housing detachment causing leakage.

Recall Class	Product Identification	Distribution	Affected Dates
II	Model No.: 4251128-02 UDI-DI: 04046964014017 Lot: 24F09G8952	3,200 units in California	February and prior

For additional information, please visit the [FDA Website](#).