

State of California—Health and Human Services Agency California Department of Public Health



Food and Drug Branch – Device Recalls

## CALIFORNIA DEVICE RECALL INFORMATION SHEET

## Cardinal Health 200 Recalls Monoject Syringe Luer-Lock Tip Soft Pack For Compatibility Issues With Syringe Infusion Pumps

| Recall Date | Product Description  | Recalling Firm                                  | Recall Reason  |
|-------------|--|---|--|
| 11/1/2023   | <b>Cardinal Health 200</b><br>Monoject 1 mL Tuberculin<br>Syringe Luer-Lock Tip Soft<br>Pack | <b>Cardinal Health 200</b><br>Waukegan,Illinois | Demonstrated reco-<br>gnition and comp-<br>atibility issues with<br>syringe infusion<br>pumps. |
| 11/1/2023   | <b>Cardinal Health 200</b><br>Monoject" 6 mL Syringe Luer-<br>Lock Tip Soft Pack             | <b>Cardinal Health 200</b><br>Waukegan,Illinois | Demonstrated reco-<br>gnition and comp-<br>atibility issues with<br>syringe infusion<br>pumps. |
| 11/1/2023   | <b>Cardinal Health 200</b><br>Monoject" 12 mL Syringe Luer-<br>Lock Tip Soft Pack            | <b>Cardinal Health 200</b><br>Waukegan,Illinois | Demonstrated reco-<br>gnition and comp-<br>atibility issues with<br>syringe infusion<br>pumps. |
| 11/1/2023   | <b>Cardinal Health 200</b><br>Monoject 20 mL Syringe Luer-<br>Lock Tip Soft Pack             | <b>Cardinal Health 200</b><br>Waukegan,Illinois | Demonstrated reco-<br>gnition and comp-<br>atibility issues with<br>syringe infusion<br>pumps  |
| 11/1/2023   | <b>Cardinal Health 200</b><br>Monoject" 35 mL Syringe Luer-<br>Lock Tip Soft Pack            | <b>Cardinal Health 200</b><br>Waukegan,Illinois | Demonstrated reco-<br>gnition and comp-<br>atibility issues with<br>syringe infusion<br>pumps  |

| Cardinal Health 200<br>11/1/2023 Monoject" 60 mL Syringe Lue<br>Lock Tip Soft Pack | <b>Cardinal Health 200</b><br>Waukegan,Illinois | Demonstrated reco-<br>gnition and comp-<br>atibility issues with<br>syringe infusion<br>pumps |
|--|---|---|
|--|---|---|

| Recall<br>Class | Product Identification   | Distribution  | Affected Dates              |
|-----------------|--|---|-----------------------------|
| I               | <b>Cardinal Health 200</b><br>Product Code: 1180100777;<br>UDI/DI: 10192253034530  | 3,113,846 Units<br>Nationwide including<br>California | September 2023<br>and prior |
| I               | <b>Cardinal Health 200</b><br>Product Code: 1180600777;<br>UDI/DI: 10192253034608  | 3,936,682 Units<br>Nationwide including<br>California | September 2023 and prior    |
| I               | <b>Cardinal Health 200</b><br>Product Code: 1181200777T;<br>UDI/DI: 10192253025811 | 3,084,300 Units<br>Nationwide including<br>California | September 2023 and prior    |
| I               | <b>Cardinal Health 200</b><br>Product Code: 1182000777;<br>UDI/DI: 10192253034677  | 4,913,520 Units<br>Nationwide including<br>California | September 2023 and prior    |
| I               | <b>Cardinal Health 200</b><br>Product Code: 1183500777;<br>UDI/DI: 10192253034691  | 1,097,048 Units<br>Nationwide including<br>California | September 2023 and prior    |
| I               | <b>Cardinal Health 200</b><br>Product Code: 1186000777T;<br>UDI/DI: 10192253025835 | 457,200 Units<br>Nationwide including<br>California   | September 2023 and prior    |

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE FDA WEBSITE

CDPH Food and Drug Branch MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435 (916) 650-6500 ● (916) 650-6650 FAX Internet Address: <u>www.cdph.ca.gov</u>

