

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Conformis Inc. restor3d Cordera Hip System. Acetabular liner, Group C 36mm Lip

Recall Date	Product Description	Recalling Firm	Recall Reason
04/15/2025	restor3d Cordera Hip System. Acetabular liner, Group C 36mm Lip Model Number: HDL-060-C22L-000101 Acetabular Liner for Cordera Hip System	Conformis Inc.	The hip cup liner that was packaged with the incorrect labeling. HDL-060-C22S- 000101 from Lot 1788258 contained a hooded liner, but it was labeled to have a standard liner.

Recall Class	Product Identification	Distribution	Affected Dates
II	Model Number: HDL-060-C22L-000101 UDI-DI: M572HDL060C22L011 Lot: 1788258	4 units in California	March and prior

For additional information, please visit the [FDA Website](#).