California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Contamac Solutions, Inc. Recalls Nutrifill Physiologic formula, Contact Lens Insertion solution

Recall Date	Product Description	Recalling Firm	Recall Reason
6/11/25	Nutrifill Physiologic formula, Contact Lens Insertion solution	Contamac Solutions, Inc.	Contact lens insertion solution may lack sterility.

Recall Class	Product Identification	Distribution	Affected Dates
II	UDI: 00850012123002/ Lot: HFH, HGC, HGI, HGN, HKF, HKK	1111 cartons in California	May 2025 and prior.

For additional information, please visit the FDA Website.