

## California Device Recall Information Sheet

### Food and Drug Branch – Device Recalls

CooperVision, Inc. Stenfilcon A Contact Lens

Recall Date	Product Description	Recalling Firm	Recall Reason
03/25/2025	<p>Brand Names: MyDay Toric, LensCrafters One Day Premium Daily Disposable Contact Lenses For Astigmatism, Pearle Vision 1 Day Premium For Astigmatism, REVEAL 1-DAY TORIC, Vision Source Reveal Toric, Voyant 1-Day Premium Toric</p> <p>Product Name: (Stenfilcon A) Contact Lens</p> <p>Model/Catalog Number: MyDay Toric</p> <p>Product Description: 54% water content, Toric, cast moulded Silicone Hydrogel, In-monomer tinted, contact lens, for Daily Disposable Wear. Blister packaged.</p> <p>Component: No</p>	CooperVision, Inc.	a limited number of lots were manufactured with an incorrect cylinder power

Recall Class	Product Identification	Distribution	Affected Dates
II	Blister Lot Numbers: 6461501582, 6461501590, 6448502235, 6472500679	1,620 lenses in California	February and prior

	<p>Carton Lot numbers:</p> <p>19346502595028 344446500117 19346502613 21455502575 21476504025 21545502813 24127502235 40702500019 21512501331 21492501653 21492501685</p>		
--	---	--	--

For additional information, please visit the [FDA Website](#).