

State of California—Health and Human Services Agency

California Department of Public Health



Food and Drug Branch - Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Covidien Recalls Mahurkar Triple Lumen Acute Dialysis Catheter For Occlusion In Tip Of Center Lumen

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|-------------|--|--------------------------------------|---|
| 1100an Dato | 1 Todast Docomption | 1.00091 | Troodii Trodooii |
| 8/23/2023 | Triple Lumen Acute Dialysis Catheter 13 cm, Curved Extensions, IC Tray REF:8888101001HP 16 cm, Curved Extensions, IC Tray REF: 8888101002HP 20 cm, Curved Extensions, IC Tray REF: 8888101003HP 24 cm, Curved Extensions, IC Tray REF: 8888101004HP 20 cm, Straight Extensions, IC Tray REF: 8888102003HP 24 cm, Straight Extensions, IC Tray REF: 8888102003HP 24 cm, Straight Extensions, IC Tray REF: 8888102004HP 13 cm, Curved Extensions, PASS Tray REF: 8888103001HP 20 cm, Curved Extensions, PASS Tray REF: 8888103002HP 24 cm, Curved Extensions, PASS Tray REF: 8888103004HP 20 cm, Straight Extensions, PASS Tray REF: 8888104003HP 24 cm, Straight Extensions, PASS Tray REF: 8888104003HP 24 cm, Straight Extensions, PASS Tray REF: 8888104004HP 20 cm, Straight Extensions, PASS Tray REF: 8888104004HP 20 cm, Straight Extensions, Kit REF: 8888340629HP 24 cm, Straight Extensions, Kit REF: 8888340637HP 16 cm, Curved Extensions REF: 8888345512HP 20 cm, Curved Extensions | COVIDIEN LP Mansfield, Massachusetts | Catheter center lumen was found to have an occlusion in the tip of the catheter; the source of the occlusion an excessive MDX, a silicone-based lubricant which coats the catheter tip. An occluded, partially or fully, and/or uncured or excessive MDX may dislodge from the catheter lead to full catheter obstruction resulting in delay to treatment and partial obstruction resulting in reduced flow, or particulate dislodgement that may result in delay to treatment, hemolysis, embolism /embolus or thrombosis/thrombus |

| | REF: 8888345520HP 13 cm, Curved Extensions, Kit REF: 8888345603HP 16 cm, Curved Extensions, Kit REF: 8888345611HP 20 cm, Curved Extensions, Kit REF: 8888345629HP 24 cm, Curved Extensions, Kit REF: 8888345637HP | | |
|-----------|--|--|------------------------|
| 8/23/2023 | Mahurkar 12 Fr Triple Lumen Acute Dialysis Catheter 20 cm, Straight Extensions, Kit REF: 8888340629 24 cm, Straight Extensions, Kit REF: 8888340637 13 cm, Curved Extensions, Kit" REF: 8888345603 16 cm, Curved Extensions, Kit" REF: 8888345611 20 cm, Curved Extensions, Kit" REF: 8888345629 24 cm, Curved Extensions, Kit REF: 8888345637 | COVIDIEN LP Mansfield, Massachusetts | Catheter tip occlusion |

| Recall Class | Product Identification | Distribution | Affected Dates |
|-----------------|--|---------------------------|----------------------------------|
| | Mahurkar 12 Fr High Pressure Triple Lumen Acute Dialysis Catheter GTIN:20884521127897, 20884521127965, 20884521127972, 20884521127989, 20884521127996, 20884521128009, 0884521128010, 20884521128047, 20884521128047, 20884521128061, 10884521128071, 20884521128071, 20884521128078, 10884521128088, 20884521128139, 20884521128139, 10884521128139, | 85327 Units Nationwide | Must include first affected date |

| 20884521128146, 0884521128163/ 20884521128160, 10884521128187 / 20884521128184 | | |
|---|----------------------------|----------------------------------|
| Mahurkar 12 Fr Triple Lumen Acute Dialysis Catheter 20884521057019, 20884521057026, 10884521006768/ 20884521006765, 10884521006782 / 20884521006789, 10884521006799, 10884521006829/ 20884521006826 | 124893 Units Nationwide | Must include first affected date |

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE FDA WEBSITE

