California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Draeger, Inc. Recalls Atlan A350XL Anesthesia Workstation

Recall Date	Product Description	Recalling Firm	Recall Reason
6/4/2026	Atlan A350XL. Model Number: 8621600. Anesthesia Workstation	Draeger, Inc. Telford, PA	The Atlan anesthesia workstation either indicated a failure of the piston ventilator before use or suffered a failure of the mechanical ventilation during use.

Recall Class	Product Identification	Distribution	Affected Dates
I	Model Number: 8621600. UDI-DI Number: 04048675556183. Serial Numbers: ASSL-0354, ASSL-0357, ASSL-0360, ASSH-0240, ASSH-0242, ASSH-0243, ASSH-0245, ASSH-0246, ASSH-0247, ASSH-0248, ASSH-0249, ASSH-0250, ASSL-0355, ASSL-0356, ASSL-0358, ASSL-0359, ASSL-0361, ASSH-0239, ASSH-0241, ASSH-0244, ASSH-0251, ASSH-0252, ASSH-0253, ASSH-0254, ASTA-0308, ASTA-0310, ASTA-0312, ASTA-0271, ASTA-0272,	394 Units Worldwide	May 2026 and prior

	ASTA-0273, ASTA-0274, ASTA-0275, ASTA-0276, ASTA-0352, ASTA-0354, ASTA-0364, ASTA-0395, ASTA-0396, ASTA-0397, ASTA-0407, ASSM-0464, ASSM-0465, ASSM-0466, ASSM-0467, ASSM-0468, ASSM-0469, ASSM-0491, ASSM-0492, ASSM-0493, ASSM-0494, ASSM-0495		

For additional information, please visit the FDA website