California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Edan Diagnostics – Telemetry Transmitter (Model iT20) Recall

Recall Date	Product Description	Recalling Firm	Recall Reason
01/22/2026	EDAN Telemetry Transmitter (Model iT20), a device that collects physiological parameters such as ECG, SpO2, and pulse rate (PR) and transmits data wirelessly to a central monitoring system.	Edan Diagnostics San Diego, CA, USA	Potential cybersecurity vulnerabilities related to improper IP address configuration identified following an FDA It Has Come to Our Attention (IHCTOA) letter.

Recall Class	Product Identification	Distribution	Affected Dates
II	**Lot Code:** N/A **Model No.:** iT20 **UDI-DI:** 06944413802261 **S/N range:** Multiple serial numbers (see full list in recall data).	1,457 Units; Nationwide including California (784 units in CA)	October 2025 and prior.

For additional information, please visit the FDA Website.