California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Encore Medical, LP Recalls EMPOWR Acetabular, Dual Mobility, Poly Bearing, ID28, OD40

Recall Date	Product Description	Recalling Firm	Recall Reason
7/21/25	Brand Name: EMPOWR Dual Mobility" Product Name: EMPOWR Acetabular, Dual Mobility, Poly Bearing, ID28, OD40 Model/Catalog Number: 952-28-40E Software Version: NA Product Description: The EMPOWR Dual Mobility" system is compatible with the EMPOWR Acetabular system, 28mm femoral heads and all hip stems Component: NA	Encore Medical, LP	Due to packaging discrepancy with the incorrect tibial inserts/implants being packaged. Issued discovered via a customer complaint.
7/21/25	Brand Name: EMPOWR Dual Mobility" Product Name: EMPOWR Acetabular, Dual Mobility, Poly Bearing, ID28, OD42 Model/Catalog Number: 952-28-42F Software Version: NA Product Description: The EMPOWR Dual Mobility" system is compatible with the EMPOWR Acetabular system, 28mm femoral heads and all hip stems Component: NA	Encore Medical, LP	Due to packaging discrepancy with the incorrect tibial inserts/implants being packaged. Issued discovered via a customer complaint.

Recall Class	Product Identification	Distribution	Affected Dates
II	Lot Code: Item: 952-28-40E GTIN: 00190446673194 Lot\s: 2224A1124A UDI-DI: 00190446673194	Distribution pattern information **2 units in California**	June 2025 and prior.
II	Lot Code: Item: 952-28-42F GTIN: 00190446673200 Lot: 2225A1124 UDI-DI: 00190446673200	Distribution pattern information2 **6 units in California**	June 2025 and prior.

For additional information, please visit the Z-2147-2025, Z-2148-2025