California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Medical Systems China Co., Ltd. Recalls Carestation 600 and 700 Series Anesthesia Systems

Recall Date	Product Description	Recalling Firm	Recall Reason
12/18/2025	GE Healthcare Carestation 600 and 700 Series Anesthesia Systems (models 610, 620, 630, 650, and 650c) are devices used to deliver anesthesia and support ventilation during surgical procedures.	GE Medical Systems China Co., Ltd. Wuxi, China	Failure to automatically switch to battery power during AC power loss may cause system shutdown and reboot, resulting in temporary disruption of ventilation and anesthetic delivery and loss of prior settings.

Recall Class	Product Identification	Distribution	Affected Dates
I	**Lot Code:** N/A; **Model No.:** Carestation 600 & 700 Series (REF 1012-9620 / 1012-9650 / 1012-9655 variants); **UDI (GTIN):** Multiple – e.g. 00195278626301, 00840682103985; **S/N range:** Not specified.	1277 Units Nationwide Including California	November 2025 and prior.

For additional information, please visit the FDA Website.