



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

GE Recalls Centricity Universal Viewer Zero Footprint Client for Potential to Display Incomplete Patient Imaging Study

Recall Date	Product Description	Recalling Firm	Recall Reason
10/15/2021	GE Centricity Universal Viewer Zero Footprint Client	GE Healthcare LLC, Waukesha, Wisconsin	There is a potential to display incomplete patient imaging study.

Recall Class	Product Identification	Distribution	Affected Dates
II	software version 6.0 SP9.x	14 Devices in California	October 2021 and prior
II	software version 6.0 SP10.x	11 Devices in California	October 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

