

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

GE Healthcare Centricity Universal Viewer, PACS-IW, Radiology, Cardiology, and PACS-IW with Universal Viewer

Recall Date	Product Description	Recalling Firm	Recall Reason
03/28/2025	GE Healthcare Centricity Universal Viewer (UV) GE Healthcare Centricity PACS-IW (PACS-IW) GE Healthcare Centricity Radiology RA60-0 (RA600) GE Healthcare Centricity Cardiology CA1000 (CA1000) GE Healthcare Centricity Centricity PACS-IW with Universal Viewer	GE Healthcare	There is a potential security vulnerability in Centricity Universal Viewer (UV), Centricity PACS-IW (PACS-IW), Centricity Radiology RA600 (RA600), & Centricity Cardiology CA1000 (CA1000) where the service login credentials are able to be identified which could allow a malicious actor with these credentials to access the system and potentially manipulate patient data.

Recall Class	Product Identification	Distribution	Affected Dates
II	72 Model Numbers: 2066908-077 – K2042VJED UDI-DI: (01)00840682103800(10)SP3209 4097001C –	California	February and prior

	<p>(01)00840682145794(10)70P103 5825832001 See additional information and product list below</p> <p>7 Model Numbers: 2042988-001 – 2052829-001 See additional information and product list below</p> <p>18 Model Numbers: 2022296-001 – MANLEGACY64 UDI: (01)00840682125260(10)20P14J 2038437016 – (01)00840682125260(10)20P14H 2038437016 See additional information and product list below</p> <p>14 Model Numbers: 2033901-001 – P00011TS UDI: (01)00840682125260(10)20P14H 2038437016 – (01)00840682125260(10)20P14 M2038437016 See additional information and product list below</p> <p>4 Model Numbers: 2049587-015 – 2104867-045 UDI: (01)00840682103800(10)6.0SP5 0412094097001D See additional information and product list below</p>		
--	--	--	--

For additional information and complete product list, please visit: [FDA Website \(Universal Viewer\)](#), [FDA Website \(PACS-IW\)](#), [FDA Website \(Radiology\)](#), [FDA Website \(Cardiology\)](#), and [FDA Website \(PACS-IW with Universal Viewer\)](#).

For further questions, please call GE Healthcare at 1-855-762-6650 and choose option 1.