

State of California—Health and Human Services Agency

California Department of Public Health



Food and Drug Branch - Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

GE Healthcare Finland Oy TruSignal Ear Sensor for Reduction of Energy

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|-------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------|
| 7/12/2023 | TruSignal Ear Sensor REF TS-E-D; Oximeter | GE Healthcare Finland Oy Finland | There is a potential reduction of energy reaching patient during defibrillation, potential contact with unintended voltage, or inaccurate measurement. |
| 7/12/2023 | TruSignal Integrated Ear Sensor with GE Connector REF TS-E2-GE; Oximeter, Ear | GE Healthcare Finland Oy Finland | There is a potential reduction of energy reaching patient during defibrillation, potential contact with unintended voltage, or inaccurate measurement. |
| 7/12/2023 | TruSignal Integrated Ear Sensor with Datex Connector REF TS-E4-GE; Monitor, physiological, patient (with arrhythmia detection or alarms); REF TS-E4-N; Oximeter | Finland Oy Finland | There is a potential reduction of energy reaching patient during defibrillation, potential contact with unintended voltage, or inaccurate measurement. |
| 7/12/2023 | TruSignal Integrated Ear Sensor with Ohmeda Connector REF TS-E4-H; Oximeter | GE Healthcare Finland Oy Finland | There is a potential reduction of energy reaching patient during defibrillation, potential contact with unintended voltage, or inaccurate measurement. |

| Recall Class | Product Identification | Distribution | Affected Dates |
|-----------------|---------------------------------------------------------------------------------------------------|----------------------------|-----------------------|
| I | TruSignal Ear Sensor GTIN 00840682103251 | 10,451 Units Nationwide | May 2023 and Prior |
| I | TruSignal Integrated Ear Sensor with GE Connector GTIN 00840682103138 | 512 Units Nationwide | May 2023 and Prior |
| I | TruSignal Integrated Ear Sensor with Datex Connector GTIN 00840682103428; 00840682103381 | 3,415 Units Nationwide | May 2023 and Prior |
| I | TruSignal Integrated Ear Sensor with Ohmeda Connector GTIN 00840682103367 | 237 Units Nationwide | May 2023 and Prior |

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE FDA WEBSITE

CDPH Food and Drug Branch
MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435
(916) 650-6500 ● (916) 650-6650 FAX
Internet Address: www.cdph.ca.gov

