



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Hamilton Recalls C6 Intensive Care Ventilator**

Recall Date	Product Description	Recalling Firm	Recall Reason
8/31/2022	<b>HAMILTON-C6 Intensive Care Ventilator</b> REF: 160021	<b>Hamilton Medical AG</b> Bonaduz, Switzerland	Ventilator status indicator board can become loose, which could lead to water ingress (disinfectants) that may lead to technical fault alarms. Multiple technical faults in a short time may force ventilators into Safety Ventilation (blower runs constantly) or Ambient Sate (inspiratory channel/expiratory valves opened; patient breaths room air unassisted) with Panel connection lost message displayed.

Recall Class	Product Identification	Distribution	Affected Dates
I	UDI-DI: 07630002808590, Serial Numbers: 5663-9046	358 Units Nationwide	June 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

