



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Integra Life Sciences recalls Codman Cerelink ICP Monitor

Recall Date	Product Description	Recalling Firm	Recall Reason
6/22/2022	Codman CereLink ICP Monitor Catalog No. 826820, 826820P. Intracranial pressure monitor	Integra LifeSciences Corp. Princeton, New Jersey	Integra has received complaints associated with ICP readings drifting to -50 mmHg (out-of-range) and may manifest an error message "sensor or extension cable failure!"

Recall Class	Product Identification	Distribution	Affected Dates
I	UDI-DI- 10381780533778; All serial numbers.	Nationwide	June 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

