

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Integra LifeSciences Corp. MediHoney[®] Wound & Burn Dressing Gel

Recall Date	Product Description	Recalling Firm	Recall Reason
02/12/2025	<p>Brand Name: MediHoney[®] Gel Product Name: MediHoney[®] Wound & Burn Dressing Gel</p> <p>Model/Catalog Number: 31815 Software Version: N/A</p> <p>Product Description: MediHoney[®] Gel with Active Leptospermum Honey is a moist dressing made of Active Leptospermum Honey in combination with gelling agents. The gelling agents act to maintain the physical integrity and viscosity of the dressing upon application. As a percentage of the overall dressing, MediHoney Gel dressing contains 80% medical grade honey by weight. The dressing helps maintain a moist environment conducive to wound healing and helps aid and support autolytic debridement. The dressing can be removed without causing trauma to the wound when a primary dressing is used.</p> <p>Component: No</p>	Integra LifeSciences Corp.	Potential that the induction seal is not completely sealed to the device tube packaging.

Recall Class	Product Identification	Distribution	Affected Dates
II	Model/Catalog Number: 31815 UDI #: 10381780486886 Lot #s: 2328, 2332 and 2333	1615 units in California	January and prior

For additional information, please visit the [FDA Website](#).