California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Intuitive Surgical, Inc. Recalls da Vinci S, Si Permanent Cautery Hook Instrument

Recall Date	Product Description	Recalling Firm	Recall Reason
3/30/2026	Brand Name: da Vinci S, Si Permanent Cautery Hook Instrument Product Name: da Vinci S, Si Permanent Cautery Hook Instrument Model/Catalog Number: 420183 Software Version: NA Product Description: EndoWrist Instruments or da Vinci S/Si Reusable instruments, including blunt and sharp endoscopic dissectors, scissors, scalpels, forceps/pick-ups, needle holders, endoscopic retractors, electrocautery and accessories are intended for endoscopic manipulation of tissue, including grasping, cutting, blunt and sharp dissection, approximation, ligation, electrocautery, suturing, and delivery and placement of microwave and cryogenic ablation probes and accessories. Please refer to attached User Manual for additional details on intended use. All da Vinci S/Si reusable	Intuitive Surgical, Inc. Sunnyvale, CA	Due to increased in complaints regarding frayed or broken pitch cables on reusable surgical instruments

	instruments with jaws use mechanical tungsten cables and pulleys to transmit motion from the input discs to the housing (on the proximal chassis), through the main shaft, and the distal instrument tip (inside the patient). Component: NA		

Recall Class	Product Identification	Distribution	Affected Dates
II	Product Code Information	22, 456 Units in California Nationwide including California	December 2025 and prior

For additional information, please visit the FDA website