



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Karl Storz Endoscopy Nasopharyngolaryngoscope for Insufficient Sterility

Recall Date	Product Description	Recalling Firm	Recall Reason
4/01/2022	KARL STORZ Slim Nasopharyngolaryngoscope 96216006US V5.0 (10-2018) 11001RD1 11001RDK1	Karl Storz Endoscopy El Segundo, California	Failure to achieve the expected six-log reduction in microorganisms following the disinfection process.

Recall Class	Product Identification	Distribution	Affected Dates
II	Nasopharyngolaryngoscope	1 unit in California	April 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

