

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Max Mobility LLC Max Mobility/Permobil Smart Drive MX2+ SpeedControl Dial Ref: MX2-3DC and Max Mobility/Permobil Smart Drive MX2+ SpeedControl Dial Ref: MX2-3DCK

Recall Date	Product Description	Recalling Firm	Recall Reason
2/11/2025	Max Mobility/Permobil Smart Drive MX2+ SpeedControl Dial Ref: MX2-3DC use in conjunction with the SmartDrive MX2+ Wheelchair Power Assist component. Max Mobility/Permobil Smart Drive MX2+ SpeedControl Dial Ref: MX2-3DCK use in conjunction with the SmartDrive MX2+ Wheelchair Power Assist component.	Max Mobility LLC	Due to faulty circuit board in speed control remote use in conjunction with a wheelchair power assist, there is a potential for loss of control of device potentially resulting in minor and major injuries.

Recall Class	Product Identification	Distribution	Affected Dates
I	All serial numbers manufactured between August 17, 2023 through November 21, 2024	11,891 units in US Nationwide including CA 8,373 units in US Nationwide including CA	February 2025 and Prior

For additional information, please visit this [FDA Website SmartDrive MX2+ SpeedControl Dial \(MX2-3DC\)](#) and this [FDA Website SmartDrive MX2+ SpeedControl Dial Kit \(MX2-3DCK\)](#).