California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Medline Industries, LP – Full Electric Basic Homecare Bed Recall

Recall Date	Product Description	Recalling Firm	Recall Reason
01/16/2026	Full Electric Basic Homecare Beds including REF MDR107003E and MDR107003E-4 (4-pack), powered adjustable beds used for patient care and positioning in homecare settings.	Medline Industries, LP Northfield, IL, USA	A potential hazard involving the hand control pendant and associated wires may cause overheating under certain conditions, posing a risk of fire.

Recall Class	Product Identification	Distribution	Affected Dates
I	**Lot Code:** All lots; **Model No.:** REF MDR107003E; REF MDR107003E-4 **UDI-DI:** 40080196321002 **S/N range:** Not specified.	138412 Units; Nationwide including California	November 2025 and prior.

For additional information, please visit the FDA Website.