

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Medline Industries, LP Recalls Convenience Kits for Various Procedures – Adult Intubation Tray & Adult Main Disp. Intub Part A, RR-Liver Trns Accessory Peds, and Convenience Kits for Various Procedures – Resp/Intubation Drawers.

Recall Date	Product Description	Recalling Firm	Recall Reason
02/19/2025	Medline Convenience Kits for Various Procedures: Adult Intubation Tray, Model Number DYNJAA254A; Adult Main Disp. Intub Part A, Model Number DYNJAA262A	Medline Industries, LP - Northfield	Kits contain recalled components due to reports of reduced skin barrier wear time. This refers to a shorter duration that a skin barrier can effectively adhere to and protect the skin before it needs to be replaced.
02/14/2025	Medline RR-Liver Trns Accessory Peds, Model Number DUNJ63397; Medical Procedure Convenience Kit	Medline Industries, LP - Northfield	Medline is recalling medical convenience kits which were assembled with specific Coloplast Foley catheters and Prostatic catheters which were subsequently recalled due to a potential sterility issue detected during testing.
02/19/2025	Medline Convenience Kits for Various Procedures: Clovis Resp/Intub Drawer 2, Model Number ACC010809; Resp/Intubation Drawer #2, Model Number ACC010813	Medline Industries, LP - Northfield	Kits contain recalled components due to reports of reduced skin barrier wear time. This refers to a shorter duration that a skin barrier can

			effectively adhere to and protect the skin before it needs to be replaced.
--	--	--	--

Recall Class	Product Identification	Distribution	Affected Dates
II	DYNJAA254A, UDI-DI 10195327686260 (each) 40195327686261 (case), Lot Number 24KBR956, 24LBA655; DYNJAA262A, UDI-DI 10195327208172 (each) 40195327208173 (case), Lot Number 24LBD397	290 Adult Intubation Tray, Model Number DYNJAA254A Shipped to California 0 Adult Main Disp. Intub Part A, Model Number DYNJAA262A shipped to California	January 2025 and Prior
II	UDI/DI 10193489736755 (ea) 40193489736756 (cs), Lot Numbers: 22HLA247, 21HLA443, 21JLA520, 22GLA335.	36 Shipped to California	January 2025 and Prior
II	ACC010809, UDI-DI 10198459012235 (each) 40198459012236 (case), Lot Number 24JDC400; ACC010813, UDI-DI 10198459012556 (each) 40198459012557 (case), Lot Number 24KDA437	186 Clovis Resp/Intub Drawer Shipped to California 480 Resp/Intubation Drawer #2 shipped to California	January 2025 and Prior

For additional information, please visit the [FDA website for Medline Convenience Kits for Procedures – Adult Intubation Tray & Adult Main Disp. Intub Part A](#), [FDA website for RR-Liver Trns Accessory Peds](#), and [FDA website for Medline Convenience Kits for Various Procedures – Resp/Intubation Drawers](#).