California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Medline Industries Recalls Thoracic Robots

Recall Date	Product Description	Recalling Firm	Recall Reason
4/1/2026	THORACIC ROBOTS, DYNJ908777B	Medline Industries Northfield, IL	Kits contain recalled Olympus biopsy valves. Olympus is investigating an increase in complaints associated with rubber fragment detachment in the slit of single use biopsy valves. Potential consequences associated with detachment of fragments from valve include foreign body in patient's tracheobronchial tree, which may require intervention for removal. Other potential consequences include inflammatory response, hypoxia, and prolongation of procedure.

Recall Class	Product Identification	Distribution	Affected Dates
II	UDI-DI 10198459269196 (EA) 40198459269197 (CS) LOTS 25EBD732 25GBG733 25IBQ613 25JBD457 25KBT352 26ABA235 26ABP735	3, 240 Units in California Nationwide including California	February 2026 and prior

For additional information, please visit the FDA website