California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Medline Industries, LP – Full Electric Lightweight Homecare Bed Recall

Recall Date	Product Description	Recalling Firm	Recall Reason
01/16/2026	MEDLINE Full Electric Lightweight Homecare Bed (REF MDR107003L), an AC-powered adjustable home-use bed designed to support patient mobility and care.	Medline Industries, LP Northfield, IL, USA	A potential hazard involving the hand control pendant and associated wires may cause overheating under certain conditions, posing a risk of fire.

Recall Class	Product Identification	Distribution	Affected Dates
I	**Lot Code:** All lots; **Model No.:** REF MDR107003L **UDI-DI:** 40080196294344 **S/N range:** Not specified.	35694 Units; Nationwide Including California	November 2025 and prior.

For additional information, please visit the FDA Website.