

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Medtronic Navigation, Inc. Medtronic PIN, 100mm and 150mm Sterile Percutaneous

Recall Date	Product Description	Recalling Firm	Recall Reason
01/27/2025	<p>Medtronic PIN, 100mm STERILE PERCUTANEOUS REF: 9733235</p> <p>Medtronic PIN, 150MM, STERILE PERCUTANEOUS REF: 97733236</p> <p>The Sterile Percutaneous Reference Pin Set is intended to be used to place and remove a pin percutaneously into bony patient anatomy for rigid attachment of a patient reference during image-guided surgeries using a Medtronic computer-assisted surgery system.</p>	Medtronic Navigation, Inc.	Due to an increase in complaint that the percutaneous reference pin would not fit into the patient reference frame or percutaneous pin adapter.

Recall Class	Product Identification	Distribution	Affected Dates
II	<p>Sterile Percutaneous Pin, 100mm</p> <p>Product Number: 9733235</p> <p>UDI-DI code: 00613994247872</p> <p>Serial/Lot Numbers: 2023071142 - 2024080530</p>	67,370 units nationwide	December and prior

	Sterile Percutaneous Pin, 150mm Product Number: 9733236 UDI-DI code: 613994247865 Serial/Lot Numbers: 2023071143 - 2024080535		
--	---	--	--

For additional information, please visit the [FDA Website](#).