

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Merit Medical Systems Inc. Recalls Rapid Response Waste Management System Kit
Containing Recalled Jiangsu Shenli Medical Production Co. Ltd. Syringes

Recall Date	Product Description	Recalling Firm	Recall Reason
8/21/2024	Rapid Response Waste Management System Kit contains recalled Jiangsu Shenli Medical Production Co. Ltd. Syringes. An endoscope and accessories is a device used to provide access, illumination, and allow observation or manipulation of body cavities, hollow organs, and canals.	Merit Medical Systems Inc. South Jordan, UT	Medical convenience kits contain plastic syringes affected by the FDA Safety Alert issued on March 19, 2024. Leaks, breakage, and other quality issues have been identified that may pose a risk to patient health.

Recall Class	Product Identification	Distribution	Affected Dates
II	UDI-DI/Catalog Number: 00884450363102/BRC-BAL-INT-SL/B	12 units in California	August 2024 and prior.

For additional information, please visit the [FDA Website](https://www.fda.gov).