

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

The Metrix Company Secure 250 mL Empty EVA Container With 2 Ports, REF 66045 and
 Secure 100 mL Empty EVA Container With 2 Ports, REF 66040

Recall Date	Product Description	Recalling Firm	Recall Reason
11/08/2024	Secure 250 mL Empty EVA Container With 2 Ports, REF 66045 Secure 100 mL Empty EVA Container With 2 Ports, REF 66040 An empty single-use pouch made of plastic intended to contain a fluid for intravenous (IV) administration.	The Metrix Company	A limited number of IV bags have been found to leak during filling.

Recall Class	Product Identification	Distribution	Affected Dates
II	UDI/DI 00812496011589 Lot Numbers: 66045-A8308, 66045-A8334, 66045-A8360, 66045-A8382, 66045-A8419 UDI/DI 00812496011541 Lot Numbers: 66040-A8314, 66040-A8368, 66040-A8417, 66040-A8482	2921 cases of 50 units worldwide and nationwide including California of the 250 ml container. 2229 cases of 50 units worldwide and nationwide including California of the 100 ml container.	October and prior.

For additional information, please visit [FDA Website-product SECURE 250 mL](#) and [FDA Website-SECURE 100 mL](#).