

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

NOA Medical Industries Inc Behavioral Health Hospital bed side rails

Recall Date	Product Description	Recalling Firm	Recall Reason
01/10/2025	NOA Medical Industries Behavioral Health Hospital bed side rails Product number 1055050GRY- BOX	NOA Medical Industries Inc	Red button used to engage side rails can break or become stuck making it difficult to engage the side rail.

Recall Class	Product Identification	Distribution	Affected Dates
II	UDI/DI 00817492012745 Serial Numbers: 2038043534 - 2038038766	22 units in California	November and prior

For additional information, please visit the [FDA Website](#).