California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

New Medical Technologies Gmbh Integra Miltex Cryosolutions Cartridges 4 pack and 10 pack

Recall Date	Product Description	Recalling Firm	Recall Reason
02/05/2025	Integra Miltex Cryosolutions Cartridges 4 Pack (23.5g N2O each) Model: C-CA-23 Item: 33517 Integra Miltex CryoSolutions Cartridges 10 Pack (23.5g N2O each) Model: C-CA-23 Item: 33516	New Medical Technologies Gmbh	During the assembly of the cartridge to its connecting body part, the valve may open prematurely resulting in partial or complete emptying of the cartridge before its intended use.

Recall Class	Product Identification	Distribution	Affected Dates
II	Model: C-CA-23 Item: 33517 UDI-DI: 10381780171966 Lots: C6741, D0122, D2560 and C6637 Model: C-CA-23 Item: 33516 UDI-DI: 10381780171959 Lots: C6637	33 units in California	December and prior

For additional information, please visit FDA Website (4 pack) and FDA Website (10 pack).