

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Olympus Corporation of the Americas Duravent Silicone Ventilation Tube

Recall Date	Product Description	Recalling Firm	Recall Reason
12/19/2025	Duravent Silicone Ventilation Tube Model Number: 240075	Olympus Corporation of the Americas	Potential that Duravent Silicone Ventilation Tubes may contain incorrect product, specifically, 1.32 mm Tiny T Tubes (model: 70240076) instead of the intended Duravent Silicone Ventilation Tubes.

Recall Class	Product Identification	Distribution	Affected Dates
II	Model Number: 240075 UDI: 00521925031883 (17)331015 (10)SM405666 Lot Number: SM405666 UDI: 00521925031883 (17)331015 (10)JB362646 Lot Number: JB362646	13 units in California	November and prior

For additional information, please visit the [FDA Website](#).