

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Olympus Corporation of the Americas Olympus MAJ-891 Forceps/Irrigation Plug

Recall Date	Product Description	Recalling Firm	Recall Reason
01/16/2025	<p>Olympus MAJ-891 Forceps/Irrigation Plug (Isolated Type)</p> <p>Model/Catalog Number: MAJ-891</p> <p>This instrument is designed to be attached to the instrument channel port of Olympus endoscopes to allow both irrigation and the use of endo-therapy accessories.</p>	Olympus Corporation of the Americas	Potential for inadvertent retention of biomaterial when the reprocessing procedure is not performed as instructed, so as not completely removing and disassembling the device.

Recall Class	Product Identification	Distribution	Affected Dates
I	<p>Model Number: MAJ-891</p> <p>UDI-DI: 04953170063114</p> <p>All lot numbers.</p>	30,140 units nationwide	December and prior

For additional information, please visit the [FDA Website](https://www.fda.gov).