



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Össur Updates Instructions for Use of Miami J Select Collar

Recall Date	Product Description	Recalling Firm	Recall Reason
11/09/2021	Miami J Select Collar and Miami J Select Collar Set Collar Item: MJS-101 Collar Set Item: MJSR-101	Össur Americas Foothill Ranch, California	Instructions for use were updated to include additional information to facilitate the selection and use of the collar device.

Recall Class	Product Identification	Distribution	Affected Dates
II	Instructions for Use Rev. 3 2018	1,539 Devices in California	November 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

